

EU Type Examination Certificate

No. 0200-NAWI-08183

C10

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Global Trade Industry.**
7 Rue Berlioze,
Belvedere, Casablanca
MOROCCO

In respect of Non-automatic weighing instrument designated C10 with variants of modules of load receptors and load cells.
Accuracy class III, single-interval, dual-range or dual-interval
Maximum capacity, Max: From 3 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n = 3000$ or 2×3000 .
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 9 pages.

Issued on **2020-03-18**
Valid until **2030-03-18**

Jens Hovgård Jensen
2020-03-18

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 120-24080.10 and ID no.: 0200-NAWI-08183

Signatory: Jens Hovgård Jensen