We help ideas meet the real world

Test Certificate / Part Certificate

No. DK0199-R60-15.01

Issued by **DELTA Danish Electronics, Light & Acoustics**
EU - Notified Body No. 0199

In accordance with EN 45501 (1992), paragraph 8.1 and 3.5.4
with fraction $p_{LC} = 0.7$
OIML R60 (Edition 2000), WELMEC 2.4. (2001)

Issued to **Hana Elektronik Tartıi Sanayi Ticaret Anonim Şirketi**
Akçaburgaz Mahallesi Alkop Sanayi Sitesi B-4, Blok No:14
Esenyurt – Istanbul
TURKEY

In respect of Shear beam load cell, with strain gauges

Type HL-H8C
Manufacturer Hana Elektronik Tartıi San. Tic. A.Ş.

Description and documentation The load cell is described and documented – including a summary of the tests - in the Annex which forms part of this test certificate.

Table 1: Essential data

Type designation		HL-H8C
Accuracy class acc. to OIML R60		C3
Maximum number of intervals	n_{LC}	3000
Maximum capacity	E_{max}	500 kg to 2500 kg
Apportionment factor	p_{LC}	0.7
Minimum verification interval	v_{min}	0.02 % E_{max}
Ratio of min LC verification interval	Y	5000
Minimum dead load output return	DR	0.0166 %E_{max}
Ratio of minimum dead load output return	Z	3000
Rated output (sensitivity)	C_n	2 mV/V
Load cell material		Alloy steel

The annex comprises 11 pages.

Issued on **2015-01-26** **Signatory: J. Hovgård**

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. (+45) 72 19 40 00
Fax (+45) 72 19 40 01
www.delta.dk
VAT No. DK 12275110

DANAK
PROD Reg. nr 7026