We help ideas meet the real world

Test Certificate / Part Certificate

No. DK0199-R60-16.03

Issued by **DELTA Danish Electronics, Light & Acoustics**
EU - Notified Body No. 0199

In accordance with EN 45501 (2015), paragraph 8.2 and 3.10.2.1
with fraction p_{LC} = 0.7
OIML R60 (Edition 2000), WELMEC 2.4. (2001)

Issued to **Hana Elektronik Tartı Sanayi ve Ticaret Anonim Şirketi**
Akçaburgaz Mahallesi Alkop Sanayi Sitesi B-4, Blok No:14
Esenyurt – Istanbul
TURKEY

In respect of Compression load cell, with strain gauges

Type HL-BM14G
Manufacturer Hana Elektronik Tartıi San. Tic. A.Ş.

Description and documentation The load cell is described and documented – including a summary of the tests - in the Annex which forms part of this test certificate.

Table 1: Essential data

Type designation		HL-BM14G
Accuracy class acc. to OIML R60		C3
Maximum number of intervals	n_{LC}	3500
Maximum capacity	E_{max}	20 t to 100 t
Apportionment factor	p_{LC}	0.7
Minimum verification interval	v_{min}	0.0625 % E_{max}
Ratio of min LC verification interval	Y	16000
Minimum dead load output return	DR	0.0132 %E_{max}
Ratio of minimum dead load output return	Z	3800
Rated output (sensitivity)	C_n	2 mV/V
Load cell material		Steel

The annex comprises 4 pages.

Issued on **2016-05-30** **Signatory: J. Hovgård**

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. (+45) 72 19 40 00
Fax (+45) 72 19 40 01
www.delta.dk
VAT No. DK 12275110

DANAK
PROD Reg. nr 7026