DANAK
PROD Reg. nr 7026

FORCE Certification

EVALUATION CERTIFICATE

No. 0200-WL-07400 Revision 1

Object name	**HP / VPI / ISC / AS / WSE**
Object type	**Non-automatic weighing indicator / weight transmitter**
Issued by	**FORCE Certification** EU - Notified Body No. 0200
In accordance with	EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017 on metrological aspects of non-automatic weighing instruments.
Fractional factor (p_i)	0.5 (refer to 3.10.2.1 of EN 45501:2015).
Issued to	**Henk Maas Weegschalen B.V.** Tuinstraat 1 AW4264 Veen Netherlands
Manufacturer	Henk Maas Weegschalen B.V.
In respect of	Non-automatic weighing indicators / weight transmitter.
Characteristics	Suitable for non-automatic weighing instruments with the following characteristics: Weighing range: Single-interval or multi-interval or multi-range. Accuracy class: III or IIII Verification scale interval: e_i = Max_i/n_i Maximum number of verification scale intervals: n_i = 10,000 per interval/range Minimum input voltage per VSI: 0.3 µV The essential characteristics are described in the annex.
Description and documentation	The non-automatic weighing indicators are described and documented in the annex to this certificate.
Remarks	Summary of tests involved: See the annex to this certificate.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission from the holder of the certificate mentioned above.

The annex comprises 64 pages.

Issued on **2024-05-27**

Jens Hovgård Jensen
2024-05-27

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 124-24644.90.10 and ID no.: 0200-WL-16953-1

Signatory: Jens Hovgård Jensen