

EU Type Examination Certificate

No. 0200-NAWI-14845 Revision 1

T2200P

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Joaquim Gabriel De Barros Rodrigues Unipessoal Lda**
Rua Pinheiro 363, Serzedo
4410-037 Vila Nova de Gaia, Porto
Portugal

In respect of Non-automatic weighing instrument designated T2200P with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III and IIII
Maximum capacity, Max: From 1 kg up to 199 500 kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 7500$ for single-interval and $n \leq 2 \times 7500$ for multi-range and multi-interval (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 / OIML R76:2006.

Note. This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

Issued on **2023-07-19**
Valid until **2033-05-30**

Jens Hovgård Jensen
2023-07-19

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 123-26808.90.10 and ID no.: 0200-NAWI-14845-2

Signatory: Jens Hovgård Jensen