DANAK
PROD Reg. nr 7026

FORCE Certification

EU Type Examination Certificate

No. 0200-NAWI-19157

KP E10 / KP S10 / KP B10

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **K&P Wagi Sp. Z.o.o.**
Paproc 134
64-300 Nowy Tomysl
Poland

In respect of Non-automatic weighing instrument designated KP E10 / KP S10 / KP B10 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III and IIII
Maximum capacity, Max: From 1 kg up to 15 000 kg
Verification scale interval: e = Max / n
Maximum number of verification scale intervals: n ≤ 6000 for single-interval and multi-range (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

Issued on **2025-05-21**
Valid until **2035-05-21**

Jens Hovgård Jensen
2025-05-21

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 125-22523.10.10.10 and ID no.: 0200-NAWI-19157-1

Signatory: Jens Hovgård Jensen

This EU type examination certificate may not be reproduced except in full, without written permission by FORCE Certification A/S