

EU Type Examination Certificate

No. 0200-NAWI-02847 Revision 3

RFC M / RFC HM / TRFC M-A / TRFC HM-A NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Kern & Sohn GmbH**
Ziegelei 1
72336 Balingen
GERMANY

In respect of Non-automatic weighing instrument designated RFC M / RFC HM / TRFC M-A / TRFC HM-A with variants of modules of load receptors and load cells.
Accuracy class III, multi-interval
Maximum capacity, Max: From 3 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n = 2 \times 3000$.
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 10 pages.

Issued on **2021-05-10**
Valid until **2027-10-02**

Jens Hovgård Jensen
2021-05-10

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 121-26435.90.10 and ID no.: 0200-NAWI-10731

Signatory: J. Hovgård Jensen