

EU Type Examination Certificate

No. 0200-NAWI-04107 Revision 4

EWJ...M / PFB...NM

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Kern & Sohn GmbH**
Ziegelei 1
72336 Balingen
GERMANY

In respect of Non-automatic weighing instrument designated EWJ...M / PFB...NM with variants of modules of load receptors and load cells.
Accuracy class II, single-interval
Maximum capacity, Max: 600 g or 6000 g
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n = 6000$.
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2021-08-12**
Valid until **2023-12-06**

Jens Hovgård Jensen
2021-08-12

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 121-30194.90.10 and ID no.: 0200-NAWI-04107-2

Signatory: Jens Hovgård Jensen