

# EU Type Examination Certificate

**No. 0200-NAWI-04955 Revision 2**

**W100 / W200 / W200IP64 / W200IP67 / WDOS / WDESK /  
WDESKLIGHT / WINOX / WTAB / W200BOX /  
W200BOXEC / ADPEW200 / WLIGHT**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **Laumas Elettronica SRL.**  
Via Primo Maggio 6  
I – 43022 Montechiarugolo (PR)  
ITALY

**In respect of**    Non-automatic weighing instrument designated W100 / W200 / WDOS / W200IP64/ W200IP67 / WDESK / WDESKLIGHT / WINOX / WTAB / W200BOX / W200BOXEC / ADPEW200 / WLIGHT with variants of modules of weight transmitters, load receptors, load cells and peripheral equipment.  
Accuracy class III and IIII  
Maximum capacity, Max: From 0.3 kg up to 999 999 kg  
Verification scale interval:  $e = \text{Max} / n$   
Maximum number of verification scale intervals:  $n \leq 10000$  for single-interval and  $n \leq 3 \times 10000$  for multi-range and multi-interval (however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of OIML R76:2006, the European Standard EN 45501:2015 and WELMEC 2.1:2001.

**Note: This certificate is a revised edition which replaces previous editions.**

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 29 pages.

**Issued on**        **2021-12-21**  
**Valid until**     **2031-03-08**

**Jens Hovgård Jensen**  
2021-12-21

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 121-31619.90.10 and ID no.: 0200-NAWI-12049-1

**Signatory: J. Hovgård Jensen**