

EU Type Examination Certificate

No. 0200-MID-03219

M2200-P02-HOP-xxxxx

**DISCONTINUOUS TOTALIZING AUTOMATIC WEIGHING INSTRUMENT
and / or
AUTOMATIC CATCHWEIGHING INSTRUMENT**

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements for the automatic weighing instruments in Directive 2014/32/EU of the European Parliament and Council of February 26, 2014 on Measuring Instruments (MID).

Issued to **Marel hf**
Austurhraun 9
210 Gardabaer
Iceland

In respect of Automatic weighing instrument designated M2200-P02-HOP-xxxxx with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class: 0.2, 0.5, 1 or 2, respective Y(a) or Y(b).
Maximum capacity, Max: ≥ 10 kg.
Verification scale interval: $e \geq 10$ g
Maximum number of verification scale intervals: $n \leq 7500$ (however dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex VIII (MI-006), chapter I, II & III of the Directive 2014/32/EU is met by the application of OIML R107-1:2007, OIML R51-1:2006, WELMEC Guide 7.2, WELMEC Guide 8.16-1, and WELMEC Guide 8.16-2.

Note: This certificate is a revised edition of DK0199.137 rev.1 extending the validation period.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

Issued on **2017-12-19**
Valid until **2027-12-19**



Signatory: J. Hovgård

FORCE Certification references:
Task no.: 117-34272 and ID no.: 0200-MID-03219