

EU Type Examination Certificate

No. 0200-MID-05152

9000 series

AUTOMATIC CATCHWEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements for the automatic weighing instruments in Directive 2014/32/EU of the European Parliament and Council of February 26, 2014 on Measuring Instruments (MID).

Issued to **Marel Limited**
Wyncolls Road
Severals Industrial Park
Colchester, CO4 9HW
United Kingdom

In respect of Automatic checkweighing / catchweighing instrument designated 9000 series with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class XIII(1) and Y(a).
Maximum capacity: $\leq 40\text{kg}$
Verification scale interval: $e \geq 1\text{ g}$.
Maximum number of verification scale intervals: $n \leq 2750$ for single interval
 $n \leq 2750$ per interval for multi-interval
(however dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex VIII (MI-006), chapter I & II of the Directive 2014/32/EU is met by the application of OIML R51-1:2006 and WELMEC Guide 7.2:2018.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2018-12-14**
Valid until **2028-12-14**

Jens Hovgård Jensen
2018-12-14

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-33891.90.30 and ID no.: 0200-MID-05152

Signatory: J. Hovgård Jensen