

EU Type-Examination Certificate

No. 0200-NAWI-05137

9010

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Marel Limited**
Wyncolls Road
Severals Industrial Park
Colchester, CO4 9HW
United Kingdom

In respect of Non-automatic weighing instrument designated 9010 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III or IIII, single-interval or multi-interval (dual)
Maximum capacity, Max: $n \times e$
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 3000$ (Class III), ≤ 1000 (Class IIII)
(however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of EN 45501:2015, OIML R76:2006. and WELMEC 2.1:2001.

The principal characteristics and examination conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2018-10-29**
Valid until **2028-10-29**

Jens Hovgård Jensen
2018-10-29

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-33891.90.10 and ID no.: 0200-NAWI-05137

Signatory: J. Hovgård Jensen