

# EU Type Examination Certificate

**No. 0200-NAWI-05307**

**APP**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**       **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**       **MURBAY ITHALAT IHRACAT ELEKTRONIK SAN. TIC.LTD.STI.**  
Kircaali Mah. Celal Bayar Cad. No: 137/A Osmangazi,  
Bursa  
TURKEY

**In respect of**   Non-automatic weighing instrument designated APP with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, dual-range or dual-interval  
Maximum capacity, Max: From 3 kg to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n = 3000$ .  
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

**Issued on**       **2019-02-01**  
**Valid until**    **2029-02-01**

**Jens Hovgård Jensen**  
2019-02-01

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 118-34735.90.15 and ID no.: 0200-NAWI-05307

**Signatory: J. Hovgård Jensen**