

EU Type Examination Certificate

No. 0200-MID-16447

QC90-2

AUTOMATIC CHECKWEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements for the automatic weighing instruments in Directive 2014/32/EU of the European Parliament and Council of February 26, 2014 on Measuring Instruments (MID).

Issued to **Newtec A/S**
Stærmosegårdsvej 18
5230 Odense M
DENMARK

In respect of Automatic checkweigher designated QC90-2 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class XIII(1)
Maximum capacity: $\leq 5500\text{g}$.
Verification scale interval: $e \geq 2\text{ g}$.
Maximum number of verification scale intervals: $n = 2750$.
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex VIII (MI-006), chapter I & II of the Directive 2014/32/EU is met by the application of OIML R51-1:2006, OIML D11:2013 section 12 & 13 with severity level 3 and WELMEC Guide 7.2.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 9 pages.

Issued on **2024-01-15**
Valid until **2034-01-15**

Jens Hovgård Jensen
2024-01-15

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 123-33931.90.10 and ID no.: 0200-MID-16447-1

Signatory: Jens Hovgård Jensen