

EU Type Examination Certificate

No. 0200-NAWI-03815

A31P

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **OHAUS Corporation**
7 Campus Drive, Suite 310
Parsippany, NJ 07054
USA

In respect of Non-automatic price computing weighing instrument designated A31P with variants of modules of load receptors and load cells. Accuracy class III, single-interval, dual-range or dual-interval. Maximum capacity, Max: From 3 kg up to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i \leq 3000$.
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

Note: This certificate is a revised edition of DK0199.438 revision 1.

The annex comprises 10 pages.

Issued on **2018-03-22**
Valid until **2024-02-11**



Signatory: J. Hovgård

FORCE Certification references:
Task no.: 117-33885 and ID no.: 0200-NAWI-03815