DANAK
PROD Reg. nr 7026

FORCE Certification

EU Type Examination Certificate

No. 0200-NAWI-07519

PC / PCS / PCP

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **R.Slonskio Enterprise "Romasas"**
Metalistu 1A,
LT-76354 Siauliai,
Lithuania

In respect of Non-automatic weighing instrument designated PC / PCS / PCP with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III and IIII
Maximum capacity, Max: From 1 kg up to 199 950 kg
Verification scale interval: e = Max / n
Maximum number of verification scale intervals: $n \leq 6000$ for single-interval and $n \leq 2 \times 3000$ for multi-range and multi-interval (however, de-pendent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, WELMEC 2.1:2001 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 14 pages.

Issued on **2019-11-15**
Valid until **2029-11-15**

Jens Hovgård Jensen
2019-11-15

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 119-35050.90.20 and ID no.: 0200-NAWI-07519

Signatory: J. Hovgård Jensen