

EU Type Examination Certificate

No. 0200-NAWI-07104

RL-HPJ / RL-EPJ

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Rice Lake Weighing Systems**
230 West Coleman Street
Rice Lake, WI 54868
USA

In respect of Non-automatic weighing instrument designated RL-HPJ / RL-EPJ with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class: III and IIII
Maximum capacity, Max: From 600 kg up to 5000 kg
Verification scale interval: $e_i = \geq 0.2$ kg
Maximum number of verification scale intervals: $n_i \leq 3000$ for single-interval, $\leq 2 \times 3000$ for multi-range or multi-interval (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 8 pages.

Issued on **2019-08-27**
Valid until **2029-08-27**

Jens Hovgård Jensen
2019-08-27

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 119-31485.90.20 and ID no.: 0200-NAWI-07103

Signatory: J. Hovgård Jensen