

EU Type Examination Certificate

No. 0200-NAWI-09305

Joerns Healthcare OXF-SCALE-LPU
NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **SR Instruments Inc.**
600 Young Street
Tonawanda, New York 14150
USA

In respect of Non-automatic weighing instrument designated as the Joerns Healthcare OXF-SCALE-LPU Digital Scale manufactured by SR Instruments, Inc. (SR Model FG4105MLF) having the following characteristics:
Accuracy class: III
Maximum capacity: 400 kg
Minimum capacity: 10 kg
Verification scale interval: $e = 0.2$ kg

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 6 pages.

Issued on **2020-11-24**
Valid until **2030-11-24**

Jens Hovgård Jensen
2020-11-24

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 120-33448.90.20 and ID no.: 0200-NAWI-09305

Signatory: Jens Hovgård Jensen