

EU Type Examination Certificate

No. 0200-NAWI-08329 Revision 1

ProCuity™ Bed Series

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Stryker Medical**
3800 E Centre Ave,
Portage, MI 49002
USA

In respect of Non-automatic weighing instrument designated ProCuity™ Bed Series with variants of modules of load receptors and load cells.
Accuracy class III, single-interval,
Maximum capacity: 250 kg
Minimum capacity: 20 kg
Verification scale interval: $e = 2$ kg
Maximum number of verification scale intervals: $n = 125$.
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2021-04-09**
Valid until **2030-11-02**

Jens Hovgård Jensen
2021-04-09

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 121-25364.90.10 and ID no.: 0200-NAWI-10590

Signatory: Jens Hovgård Jensen